





REPLACEMENT COVERS

RANGE OF POSTURAL AIDS FOR BED

RANGE	ITEM CODE	DESIGNATION	WASHING CONDITIONS
 <p>POSTURAL EQUIPMENTS MADE OF FOAM</p>	P901HW	Cover for Integral Heel Support	
	P902HW	Cover for Heel Pad	
	P904HW	Cover for Half-Heel Pad	
	P906HW	Cover for Heel Pad with abduction wedge	
	P903HW	Cover for Heel Support Boot	
	P905HW	Cover for Semi-Fowler Retraction Wedge	
	P910HW	Cover for 30° Positioning Wedge	
	P910LGHW	Cover for Half 30° Positioning Wedge	
	P910LDHW	Cover for Half 30° Positioning Wedge	
	P912HW	Cover for Positioning Wedge for the Upper Limbs (Item code P912L1HW)	
	P915HW	Cover for Positioning Wedge for the Upper Limbs (Item code P915L1HW)	
	P911HW	Cover for Abduction Wedge	
	P913HW	Cover for Occipital System	
	P914HW	Cover for Universal cushion	
 <p>POSTURAL EQUIPMENTS MADE WITH MICRO-BEADS OR SYNTHETIC FIBRE</p>	P9700HW	Cover for Universal Cushion «Small»	
	P9701HW	Cover for Universal Cushion «Standard»	
	P9702HW	Cover for Cylindrical Cushion	
	P9703HW	Cover for Abduction Block for Knees	
	P9704HW	Cover for Ring Cushion for Neck	
	P9705HW	Cover for Half-Moon Cushion	
	P9706HW	Cover for Removable Back Cushion for Half-Moon Cushion	
	P9707HW	Cover for Lateral Positioning Cushion	
	P9709HW	Cover for Cross pad	
	P9710HW	Cover for Delta cushion	
	P9711HW	Cover for Hips abduction wedge	
	P9712HW	Cover for Positioning wedge with front / back slope	
	P9713HW	Cover for Occipital system	
	P9714HW	Cover for Universal Cushion «Extra small»	



Disinfect using cold-application spray products
(quaternary ammoniums, glutaraldehydes, etc).

GOOD TO KNOW... ON THE COVER'S LABEL

Item code	LOT TU5201
Composition	MD 11 CE 26709712 99709712
Cleaning instructions	REF P9713HW 80% PES, 20% PU 80% PES, 20% PU Washing instructions: 95°C, 1000 RPM, 15 min, 15°C, 1000 RPM, 15 min, 15°C, 1000 RPM, 15 min
Free area	LA PRESENCE PROVENÇON DE L'ENCAPTE P9713HW ASSOCIÉE À UN SCHEMA P9713HW P9713HW P9713HW P9713HW P9713HW P9713HW L'ARTICLE A LA COMPOSITION PROPOSÉE DANS LE FICHIER AVEC LE PRESENT DOCUMENT EST CONFORME À LA NORME EN 13060-1